



## HPT Sensor Calibration Certificate

Part Number: 210091  
 Maximum Pressure (psi): 100

Serial Number: XD31080A  
 Calibration date: July 10, 2018  
 Technician: SPB

### Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.947E-1	-1.381E+0

### Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.680	16.126	14.660	0.020
20 PSI	21.310	22.791	21.289	0.021
30 PSI	30.790	32.346	30.794	-0.004
40 PSI	40.470	42.095	40.491	-0.021
50 PSI	50.450	52.135	50.478	-0.028
60 PSI	60.030	61.754	60.046	-0.016
70 PSI	70.080	71.837	70.075	0.005
80 PSI	79.780	81.553	79.740	0.040